Operator Tips

Logbook Entries

The precise format for entering data in the Field Ratemeter logbook is flexible. However the following example illustrates the basic information that should be included. Annually you will receive a newly calibrated field counter, along with its logbook and check source. We will provide a prepaid, preaddressed shipping label for returning your current counter, logbook and check source.

05 August 2010

Background CPM: 194 (monthly avg = 209 CPM)
This is within 20% of monthly avg background.

Beta Check Source CPM:

1800 – 209 = 1591 Net CPM
This is within lower/upper control limits.

Radney Jones

Notes:

You may wish to write the Upper/Lower Control Limits on a page in the Logbook – so you don't need to refer to the Certificate sheet each time.

Please refer to Standard Operating Procedure FMM/SOP-10 or RadNet/SOP-12 to review procedures for using the counter.